



Automated External Defibrillator

ADDENDUM NO. 1

Date: January 29, 2024

RE: Automated External Defibrillator

BID DATE AND TIME:

Tuesday, February 6, 2024, 2:00 P.M.

Location for the bid opening remains the Shelby County Manager's Office, located at 200 West College Street, Room 123, Columbiana, Alabama, 35051

* * * * *

Clarification:

Q1. Warranty length

A1. We did not specify a warranty length. I would recommend using the manufacturer's warranty.

Q2. Life of pads/batteries-this can be anywhere from 2-5 years.

A2. The bid specifies two-year minimum shelf life.

Q3. Compatibility/standardization with local Fire/EMS and area hospitals-since we are a local company, we know what defibrillator models are used by all local EMS transport and all area hospitals. There is a definite benefit to having the same technology that they use.

A3. The previous years we have received the following models: Rugged - Phillips Heartstart Onsite and Non-rugged - Heartsine 360.

Q4. IP rating-will any of the AEDs be stored outside (parks, pool, etc.)?

A4. All units will either be in a building or in a vehicle.